

Kliinisen tutkimuksen keskus (CRC)
Palaveri 17.10.2018 klo 8:30
CRC:n neuvotteluhuone

Asiat:

1. Uudet työntekijät
2. Työpaikkaselvityskäynti (18.10)
3. Sentrifugien roottorit
4. Zeiss mikroskooppi
5. Revco3 syväjäähakastin
6. Muut asiat

Mistä keskusteltiin:

- 1.
2. Terveystalolta ilmoitettiin, että kyselyjen vastausprosentti on ollut hyvä ja kyselyihin oltiin vastattu kattavasti. Käytiin läpi Terveystalon työpaikkakäynnin (18.10.) ohjelma. Terveystalon työpaikkaselvityksen palautetilaisuus on kaikille avoin ja se pidetään 1.11.2018 klo 13-14:30 salissa F202. Ergonomiaan liittyvä infotilaisuus pidetään eri päivänä ja siitä ilmoitetaan erikseen.
3. Ilmoitettiin, että CRC:n sentrifugien roottorit tullaan lähiaikoina tarkistamaan (tämä on säännöllinen tarkistus). Vialliset roottorit tullaan viemään pois. Samalla kartoitetaan, että onko joistakin roottoreista puute.
4. ZEISS mikroskoopin ohjelma päivitetään ja tietokone vaihdetaan uuteen. Tämä takia käyttäjien pitää kopioida omat tiedostot talteen. Tästä on laitettu kaksi erillistä sähköpostia CRC sähköpostilistalle. Ohjelma päivitys ja tietokoneen vaihto suoritetaan loka-marraskuun vaihteessa. Tarkempi ajankohta ilmoitetaan heti kun se on tiedossa.
5. Rikkoutuneen syväjäähäpakkastimen (Revco3) tilanne oli edelleen epäselvä. Lupaa korjaukselle tai uuden ostamiselle ei oltu saatu. Koska syväjäähäpakkastin, jonne CRC:n pakkastimen näytteitä oli siirretty tilapäissäilytykseen, alkoi oireilla (lämpötila nousi), päätettiin perustaa ”tarkistusrinki”. Tarkoituksena oli, että jokainen CRC:n tutkimusryhmä vuorollaan käy tarkistamassa oireilevan syväjäähäpakkastimen lämpötilan. Tarkistusajankohdista on lähetetty erillinen aikataulu CRC:n sähköpostilistalle.
6. 15.6.2018 suoritettiin CRC:n tiloissa vaarojen tunnistaminen ja riskien arviointi, jossa oli mukana edustajia COOR:lta ja PPSHP:ltä. Ko arvioinnin palaute käytiin läpi. Laura Kytövaara kertoi uudesta näytekeräyksestä. Näytteitä toimitetaan ensin kylmähuoneeseen (360B), josta tutkimusryhmä siirtää näytteet omiin tiloihin. Isotoopeilla työskentely CRC:n tiloissa melko varmasti loppuu vuoden 2018 aikana.

What was discussed:

- 1.
2. Terveystalo informed that the response rate of the questionnaires was good and that the questionnaires were answered very well. We went through the Terveystalo workplace visit program (18.10.). The feedback session of Terveystalo's workplace survey is open to everyone and will be held on 1.11. at 13-14:30 in Hall F202. Ergonomic information is held on a different day and is reported separately.
3. It was announced that the CRC's centrifuge rotors will be checked in the near future (this is a regular review). Defective rotors will be removed. At the same time, it will be checked whether some of the rotors are lacking.
4. The ZEISS microscope software is updated and the computer is replaced with a new one. This is why users need to back up their own files. There have been two separate e-mails on this sent to CRC mailing list. The software update and computer replacement will be performed in late October and early November. A more accurate time is announced as soon as it is known.
5. The situation of the broken deep freezer (Revco3) was still unclear. No repair or new purchase was granted. Since the deep-freezer, where the CRC freezer samples had been moved to temporary storage, began to have problems with temperature, it was decided to set up a "checklist". The purpose was for each CRC research group to check the temperature of the deep freezer in turn. A separate schedule for this has been sent to the CRC mailing list.

6. On 15.6.2018 CRC's premises were identified for hazard identification and risk assessment involving representatives of COOR and PPSHP. The feedback of this was presented. Laura Kytövuori told us about a new sample collection. The samples are first delivered to the cold room (360B), from which the research team transfer the samples to their own premises.
Working with isotopes at CRC's premises will pretty much end in 2018.